

CORRECTION

Open Access



Correction to: Validity and reliability of serratus anterior hand held dynamometry

Jos IJspeert^{1*}, Hans C. J. W. Kerstens^{2,4}, Renske M. J. Janssen¹, Alexander C. H. Geurts¹, Nens van Alfen³ and Jan T. Groothuis¹

Correction to: BMC Musculoskelet Disord 2019 20:360
<https://doi.org/10.1186/s12891-019-2741-7>

Following publication of the original article [1], the authors reported that the headers in Table 2 in their paper were omitted. The original article [1] has been updated.

The correct Table 2 is shown below.

Table 2 test-retest and interrater reproducibility of serratus anterior hand held dynamometer strenght testing

	Test (N) mean (SD)	Retest (N) mean (SD)	Diff test-retest (N) mean (SD)	95% LoA (N)	ICC3.1 _{agreement} (95% CI)	SEM _{agreement} (N)	SDC _{agreement} (N)
<i>test-retest</i>							
Pos A	383.88 (77.65)	404.92 (84.96)	- 21.05 (60.22)	- 139.08; 96.98	0.712 (0.420; 0.871)	44,10	122,40
Pos B	314.43 (75.1)	322.34 (82.34)	-7.90 (65.90)	- 137.06; 121.26	0.658 (0.324; 0.846)	45,80	127,00
Pos C	351,54 (100.36)	376.16 (93.55)	-24.62 (59.37)	- 140.99; 91.75	0.794 (0.490; 0.916)	44,52	123,40
	<i>Tester JI</i> mean (SD)	<i>Tester HK</i> mean (SD)	<i>Diff Tester JI VS HK</i> mean (SD)				
<i>Interrater</i>							
Pos A	383.88 (77.65)	340.44 (82.95)	43.43* (92.74)	- 138.34; 225.20	0.794; (0.552; 0.912)	55,80	196,80
Pos B	314.43 (75.10)	264.21 (61.76)	50.21* (78.85)	- 104.34; 204.76	0.277 (-0.089; 0.605)	55,80	154,50
Pos C	351.54 (100.36)	265.10 (40,0)	86.48* (86.21)	-82.49; 255.45	0.226 (-0.107; 559)	85,31	236,47

Legend: ICC 3.1 intraclass correlation coefficient model 3.1, 95%CI 95% confidence interval, LL Lower limit, UL upper limit
 N Newton, SD Standard deviation, Diff Difference, LoA limits of agreement, ICC Intraclass correlation coefficient, CI Confidence interval, SEM Standard error of measurement, SDC Smallest detectable change, % percentage

*: $p < 0.001$

Author details

¹Department of Rehabilitation, Donders Institute for Brain, Cognition and Behaviour, Radboud University Medical Center, Nijmegen, the Netherlands.

²Department of paramedical studies, HAN University of Applied Sciences, Nijmegen, The Netherlands. ³Department of Neurology and Clinical

Neurophysiology, Donders Institute for Brain, Cognition and Behaviour, Radboud University Medical Center, Nijmegen, The Netherlands. ⁴IQ Healthcare, Radboud Institute for Health Sciences, Radboud University Medical Center, Nijmegen, The Netherlands.

Published online: 17 September 2019

Reference

1. IJspeert J, et al. Validity and reliability of serratus anterior hand held dynamometry. *BMC Musculoskelet Disord*. 2019;20:360. <https://doi.org/10.1186/s12891-019-2741-7>.

* Correspondence: jos.ijspeert@radboudumc.nl

¹Department of Rehabilitation, Donders Institute for Brain, Cognition and Behaviour, Radboud University Medical Center, Nijmegen, the Netherlands
 Full list of author information is available at the end of the article

